

15/1/21

Requested Janardhanan to provide quotation for fixing the timer board and to change the tube head

Avarell

31/1/22

Rotational parking sensor faulty replaced to timing sensor by Mr. Yogesh. Sensor given for repair N3800

WJ

22/3/22.

Room No: 8 : Replaced New tube head and

Replaced (Temp) Timer Board (Cube Temp)

To Rectify : (Old board already existing to be serviced)

22/3/22.

CONDEMNED TUBE HEAD of RNO 6

Room No. 9 & 6.

(Error 45) • tube head of Room No. 9 replaced for RNO. 6 (temporary)

To Rectify : New tube head to NO: 6 to be replaced by cube.



# SERVICE REPORT

## CUBE DENTAL EQUIPMENTS

Plot No.3, Ekambaram Street, Udhaya Nagar Extn., (Near P.J.N Mahal), Porur, Chennai - 600 116.  
 Email : cubedentalequipments@gmail.com Web : www.cubedentalequipments.com

Service Report No. **028**

Date : **22/3/2022**

CUSTOMER NAME : <u>Ragas Dental College</u>		Equipment Name : <u>Xmind DC</u>
ADDRESS : <u>Dept. of Oral Medicine &amp; Radiology.</u> <u>Room No: 5</u>		Serial No : _____
CONTACT NO : _____		Accessories Serial No : _____
CITY : _____	PIN : _____	Warranty : Under/Outside

**Fault Description :** Xray Shows Error 20.  
Timer Board Need to be Serviced  
(72814)

**Fault Solution :** 72814 Taken for service. And we  
given stand by timer board.

**Remarks :** Now All three Rooms (8, 6, 5)  
units are running fine. Taken  
almost 4 exposures from each Room.

Service Engineer Sign : <u>[Signature]</u>	The Equipment has been repaired to my satisfaction Customer Sign : <u>[Signature]</u>
--	--

Parts Replaced	Qty	Amount	
			<b>Payment Details :</b> Cash / Cheque No : _____ Bank : _____ Date : _____ Amount : _____ Service Engineer Sign : _____
<b>Service Charges</b>			
<b>TOTAL</b>			



# SERVICE REPORT

Mob : 87 54 57 58 56

## CUBE DENTAL EQUIPMENTS

Plot No.3, Ekambaram Street, Udhaya Nagar Extn., (Near PJN Mahal), Porur, Chennai - 600 116.  
Email : cubedentalequipments@gmail.com Web : www.cubedentalequipments.com

Service Report No. 027

Date: 22/3/2022

CUSTOMER NAME: Rajas Dental College  
 ADDRESS: Dept. of Oral Medicine & Radiology  
(Room No: 6)  
 CONTACT NO: \_\_\_\_\_  
 CITY \_\_\_\_\_ PIN \_\_\_\_\_

Equipment Name : Xmind DC  
 Serial No : \_\_\_\_\_  
 Accessories Serial No : \_\_\_\_\_  
 Warranty : Under/Outside

Fault Description : found E.45 : Need to  
replace New tube head. Timer  
Board is fine.

Fault Solution : As Room No: 8 tube head working  
Sometime a rattling trouble in some times.  
Time being Room No: 8 tube head  
connected in Room No: 6. As soon as  
New tube head comes will replace.

Remarks : Room no: 6 faulty tube head  
D32057 / 64714 Need to be disposed.

Service Engineer Sign: [Signature]

The Equipment has been repaired to my satisfaction  
 Customer Sign: [Signature]

Parts Replaced	Qty	Amount
<b>Service Charges</b>		
<b>TOTAL</b>		

**Payment Details :**  
 Cash / Cheque No : \_\_\_\_\_  
 Bank : \_\_\_\_\_  
 Date : \_\_\_\_\_  
 Amount : \_\_\_\_\_  
 Service Engineer Sign : \_\_\_\_\_



# CUBE DENTAL EQUIPMENTS

Plot No.3, Ekambaram Street, Udhaya Nagar Extn., (Near PJN Mahal), Porur, Chennai - 600 116.  
Email : cubedentalequipments@gmail.com Web : www.cubedentalequipments.com

Service Report No. 026

Date : 22/3/2022

CUSTOMER NAME : Ragas Dental College  
 ADDRESS : Dept. of Oral Medicine & Radiology  
(Room No: 8)  
 CONTACT NO : \_\_\_\_\_  
 CITY : \_\_\_\_\_ PIN : \_\_\_\_\_

Equipment Name : Xmind DC  
 Serial No : \_\_\_\_\_  
 Accessories Serial No : \_\_\_\_\_  
 Warranty : Under/Outside

Fault Description : Found Tube head Noise with Error E001.  
(72892)  
Some times E007. Its board related issue.

Fault Solution : Replaced New tube head & Timer Board <sup>Temp Porary</sup>  
(72892)  
S.No: XD 577595 tube: 96541

Board 72892 taken for Service.  
 Remarks : Earthing is good. So we preferred to change new tube head in this room first.

Service Engineer Sign :

The Equipment has been repaired to my satisfaction  
 Customer Sign :

Parts Replaced	Qty	Amount	Payment Details :
			Bank :
			Date :
			Amount :
Service Charges			Service Engineer Sign :
TOTAL			

22/3/22

Room NO: (5)

Board Problem


To Rectify: Replaced with (stand by cube)  
Service the old board

Taken out for service

→ 2 timer Board taken out for ~~service~~  
service by cube on 22/3/22  
(72892, 72814)  
S. No.

To be disposed like head of RNO: (6)  
1232057

N. Kumar  
22/3/22

  
(Mr. Murali)

8/4/22

## Service of Online UPS:

Ups is almost Running condition. Battery charging is proper. Dc voltage and current 204 VDC / 6A

Ups backup power is more than  $\frac{1}{2}$  hour or less.

Back-up time -  $\frac{1}{2}$  hour.

Earthing is good. [0.025 v]

Ups Cooling fan has need to be Replaced.

4 No's - Battery is in weak condition.



No problem as of now. Bless of this.

N. K. Sharma  
8/4/22

18/4/22

Neutral and Earth circuit are checked  
to be tested after opening the Junction boxes  
and checked

One complete day is spent

Proper earthing needs to be checked

Earthing not proper in (8), OPA rooms and computer room

(Seen by Mr. Dharaman - electrician)

N. K. Sharma  
18/4/22

Gi. Dharaman

2/04/22

Neutral and earthing link found out

The stabilizer in room no ② was found to be at fault linking the earth with neutral

The stabilizer was removed and kept in the factory room

The earthing was wiring 05 in all the COPA rooms and OPG rooms

G. Dharamach

22/4/2022

Machine showing E5 Problem on the Power Supply board (L 1910)

Now New power supply board is replaced and the machine is working well after the replacement of the new board.

Advice: Don't use generator power for operating the machine.

Two images are taken Machine working well.

K. Murugavel K. [Signature]

8754474614

26/04/22.

E4 Error due to generator

Board, Now that PCB has been checked and its started working took 2 to 3 images and also checked the earthing

25/04/2022  
[Signature]  
[Signature]



CUBE DENTAL EQUIPMENTS

Plot No.3, Ekambaram Street, Udhaya Nagar Extn., (Near PJN Mahal), Porur, Chennai - 600 116.

Email : cubedentalequipments@gmail.com Web : www.cubedentalequipments.com

Service Report No. 511

Date : 14/9/2022

CUSTOMER NAME : Ragas dental college  
 ADDRESS : Radiology department  
Urandi  
 CONTACT NO : 98416 98865  
 CITY Chennai PIN 600119

Equipment Name : Xmind  
para D+  
 Serial No :  
 Accessories Serial No :  
 Warranty :  Under/Outside

Fault Description : Last time Xmind Xmind para D+ machine was having some issue. Our engineer Mr. Yogam come + check problem in then find problem in the filament board (N3200) now we taken new board replaced its  
 Fault Solution : not done

Again its showing Error E10 its related to digital the board first we get the board to check cleaning purpose then we start on the issue.  
 Remarks :

[Signature]  
 Service Engineer Sign :

The Equipment has been repaired to my satisfaction  
[Signature]  
 Customer Sign :

Parts Replaced	Qty	Amount
Service Charges		
TOTAL		

Payment Details :  
 Cash / Cheque No :  
 Bank :  
 Date :  
 Amount :  
 Service Engineer Sign :



4/5/2022

Error E5 problem on the P1910 board which was replaced on 22/4/22. So New P1910 board was replaced along with L1500 board as this could be the problem for demaying the new P1910 board. Though <sup>L1500</sup> ~~L1500~~ board was working fine, it was replaced.

K. Mungavel  
K. M

12/09/22

In room no : 5 E09 error occurs due to Control Panel Problem. That panel taken to office.

In room no : 6 E33 occurs sometime and abnormal sound created from the tube head. We will raise the complaint against the new replacement.

In room no ② - Stacker not working DL  
(Premkumar)  
88: 87544745

12/09/22

In OPA/ Latral cyph - Service attached by Mr. Janarthanan. We checked ~~parts~~ xmind para Dt machine 2r's his problem in I/O digital board. Need to get board for prob. check. propose man we finished in service.

le. fu  
87544745



# RAGAS DENTAL COLLEGE & HOSPITAL

(Unit of Ragas Educational Society)  
Recognized by the Dental Council of India, New Delhi  
Affiliated to The Tamilnadu Dr. M.G.R. Medical University, Chennai  
2/102, East Coast Road, Uthandi, Chennai - 600 119, INDIA.  
Tele : (044) 24530002, 24530003 - 06. Principal (Dir) 24530001  
E-mail : principal@rdc.ac.in, web : www.rdc.ac.in

Sl. No.	Date	Particulars	Remarks	Signature
15	12-05-18	Limited space for patient's seating in Ground floor	10 new chairs were procured for patient seating.	Jan
16	29-05-18	External Cracks on Administrative block.	Repair of cracks in Administrative block.	f
17	7-06-18	Uneven ground in parking area.	levelling and maintainance of ground in the parking area.	f
18	25-06-18	Requirements of Ac in the Administrative block.	Installation of 2 Ac's in the Administrative block.	ly
19	04-07-18	difficulty in closing the main gate.	Maintainance of main gate (locking).	
YEAR (2017-2018)				At Uthandi

PRINCIPAL  
RAGAS DENTAL COLLEGE AND HOSPITAL  
UTHANDI, CHENNAI-600 119.



# RAGAS DENTAL COLLEGE & HOSPITAL

(Unit of Ragas Educational Society)  
Recognized by the Dental Council of India, New Delhi  
Affiliated to The Tamilnadu Dr. M.G.R. Medical University, Chennai  
2/102, East Coast Road, Uthandi, Chennai - 600 119, INDIA.  
Tele : (044) 24530002, 24530003 - 06. Principal (Dir) 24530001  
E-mail : principal@rdc.ac.in, web : www.rdc.ac.in

	DATE			
8	18/01/2019	Requirement of additional light in boys hostel	New lights are installed	f
9	25/02/2019	Requirement of more chairs for visitors in administration block	New chairs are placed in administration block	f
10	08/04/2019	Trimming of plants to be done	Trimming and planting work is done	f
11	25/04/2019	painting of front area to be done	painting is done	f
12	27/05/2019	Repair of pavement to the classroom	pavement is repaired	f

YEAR (2018-2019)

PRINCIPAL  
RAGAS DENTAL COLLEGE AND HOSPITAL  
UTHANDI, CHENNAI-600 119.



# **RAGAS DENTAL COLLEGE & HOSPITAL**

(Unit of Ragas Educational Society)  
Recognized by the Dental Council of India, New Delhi  
Affiliated to The Tamilnadu Dr. M.G.R. Medical University, Chennai  
2/102, East Coast Road, Uthandi, Chennai - 600 119. INDIA.  
Tele : (044) 24530002, 24530003 - 06. Principal (Dir) 24530001  
E-mail : [principal@rdc.ac.in](mailto:principal@rdc.ac.in), web : [www.rdc.ac.in](http://www.rdc.ac.in)

---